Fold Fold Fold

91 mm | 101 mm | 101 mm | 91 mm

178 mm

Also available from the Lumex line of Patient Aids.....

Tambien disponible de la llinea de Asistencias para Pacientes de Lumex

Low Profile Quad Canes
Bastones con Base Cuadrada de Perfil Bajo
Item 6140A shown here

Three-in-One Commodes
Inodoro de Tres-en-Uno
Item 7103A-4 shown here

Maxi-Drain Transfer Benches
Bancos de Traslaerr de Drenaje Maximo
Item 7921A shown here

Lumex Mobility Lights
Item(s) L780X (K, R, W, B)

Walker Tray
Bandeja Para Andador
Item 6030904

Walker Basket
Canasta de Andador
Item 6039904

Glide Skis
Esqui Deslizante
Item 6032104

Outside

Garantia Limitada

La información de garantía siguiente es un SUMARIO de la Garantía del Producto completa provisto de GF Health Products, Inc. Para la garantía completa, por favor de visitar a la página de web www.grahamfield.com

GF Health Products, Inc. ("GF") garantía al comprador original o arrendatario (cual quiera is el "Cliente") solamente, que va reemplazar o reparar componentes, a la decisión solo de GF, cuales son defectos en material o calidad utilizado normalmente y con servicio. La garantía es para partes solamente y no incluye fuerza laboral o el costo de transporte. Todas garantías con condicionales por el uso apropiado del producto estrictamente en acuerdo con las instrucciones aplicables de GF. Producto próvido a usuarios como réntales debes ser serviciados por el proveedor y inspeccionado antes de entrega. Cada usuario debe ser preparado en la operación y seguridad de este producto antes de uso. Dentro las reglas normales en este documento, la estructura tiene garantía por tres (3) anos y los componentes de uso normal como puntas, ruedas y manijas tienen garantía por tres (3) meses. El periodo de garantía se comencia en la fecha de transporte al Cliente a menos que haya una fecha de expiración en el componente en cual la garantía se expira en la periodo de garantía o la fecha de expiración.

OBTENIENDO SERVICIO DE GARANTIA
La garantía limitada solamente aplica a defectos reportados al Distribuidor de quien el Cliente compro el producto dentro el periodo de garantía aplicable. Si no hay Distribuidor, debes llamar a GF directamente llamando 770-368-4700, mandando fax a (770)-368-2386 o por email a cs@grahamfield.com. Direcciones específicos se van a dar por el Representante de Servicio De Clientes.

EXCLUSIONES
La garantía no cubre y GF será responsable por lo siguiente: 1) Defectos, daño o otras condiciones causadas por mal uso, abuso, negligencia, alteración, accidente, daño de transporte o falta de buscar y obtener repara o repuesto en manera puntual; 2) Productos no instalados, utilizados or mantenidos correctamente; y 3) Cargos de Labor o transporte. Otras exclusiones se aplican.

GARANTIA ENTERO, REMEDIO EXCLUSIVO Y DESCARGO DE DAÑOS CONSIGUIENTE
ESTA GARANTIA ES LA UNICA GARANTIA DE GF Y ES EN VEZ DE TODAS OTRAS GARANTIAS, EXPRESIDAS O IMPLICITADAS.

Para la garantía completa por este producto, por favor de visitar a la página www.grahamfield.com

LUMEX®

4mm Hole Drilled

- Improves mobility and allows for greater independence for the user on the go
- Lightweight design makes rollator easy to use, fold and store for convenient transport
- Padded seat and backrest provide additional comfort
- Ergonomic hand grips and hand brakes allow user to operate with ease
- Complete with pouch for storage of personal items
- 6" wheels allow for ease of movement
- 300 lb weight capacity - EVENLY DISTRIBUTED

- Mejora movilidad y provee mas independencia al usuario siempre en la calle
- Diseño ligero hace que el rodador es fácil de usar, doblar y almacenar para transporte conveniente
- Asiento y respaldo acolchado provee mas comodidad
- Agarres de mano ergonómicas y frenos de mano permite a usuario operarlo con facilidad
- Viene completo con bolsa para almacenaje de cosas personales
- Ruedas de 6" permitan movimiento facil
- Capacidad de peso de 300 lb - DISTRIBUIDO UNIFORMEMENTE

1.770.368.4700
Information contained herein is subject to change.
The most current and complete product information can be found on our website.
Información aquí es sujeto a cambio.
La información de producto mas actual y completo se puede encontrar en nuestra páginas web.
www.grahamfield.com
Follow us on

© 2016, GF Health Products, Inc. All Rights Reserved.
Lumex is a trademark of GF Health Products, Inc.
GF Health Products, Inc. is an ISO 13485 Certified Company.

GFI1600007Rev16

Back Cover

6mm 5mm

LUMEX®

Walkabout Essentials Steel
4-Wheel Rollator
Item RJ5000

User Instructions and Warranty Information
Important: Read Before Use
Instrucciones e Informacion de la Garantía
Importante: Lea Antes de Uso

grahamfield

Front Cover

Limited Warranty

The following warranty information is a SUMMARY of the complete Product Warranty provided by GF Health Products, Inc. For the complete warranty, please visit our website at www.grahamfield.com.

GF Health Products, Inc. ("GF") warrants to the original retail purchaser or lessee (either is the "Customer") only, that it will replace or repair components, at GF's sole discretion, that are defective in material or workmanship under normal use and service. The warranty is for parts only and does not include labor or costs of shipping. All warranties are conditioned upon the proper use of the product strictly in accordance with applicable GF instructions. Product provided to users as rentals must be serviced by the provider and inspected prior to delivery. Each user must be properly trained on the operation and safety of this product prior to use. Within the guidelines set forth in this document, the frame is warranted for three (3) years and the normal wear components such as tips, wheels/castors, and handgrips are warranted for three (3) months. The applicable warranty period shall commence from date of shipment to the Customer, unless there is an expiration date on the component in which case the warranty shall expire on the earlier of the warranty period or the expiration date.

OBTAINING WARRANTY SERVICE
This limited warranty shall only apply to defects that are reported to the Dealer from whom the Customer purchased the product within the applicable warranty period. If there is not a Dealer, you must contact GF directly by calling 770-368-4700, sending a fax request to 770-368-2386 or by e-mailing a request to cs@grahamfield.com. Specific directions will be provided by the Customer Service Representative.

EXCLUSIONS
The warranty does not cover and GF shall not be liable for the following: 1) Defects, damage, or other conditions caused by misuse, abuse, negligence, alteration, accident, freight damage, tampering or failure to seek and obtain repair or replacement in a timely manner; 2) Products not installed, used, or properly cleaned and maintained; and 3) Labor or shipping charges. Other exclusions apply.

ENTIRE WARRANTY, EXCLUSIVE REMEDY AND CONSEQUENTIAL DAMAGES DISCLAIMER
THIS WARRANTY IS GF'S ONLY WARRANTY AND IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED.

For the complete warranty for this product, please visit our website at www.grahamfield.com.

Fold Fold Fold

Instructions for Use – Please Read All Instructions Before Use.

Please consult a healthcare professional for the proper adjustment.

Assembling the Rollator

1. Remove Rollator and other components from carton.
2. Unfold frame; spread front and back legs apart.
3. Push down locking mechanism to secure legs in place.
   NOTE: Ensure the side crossbars (locking mechanism) lock down.
4. Insert handgrip handle bars into frame.
   NOTE: After installation, make sure brake cable is secured to frame with rollator clip.
5. Secure positioning of handle bars with adjustment knob.
6. Install backrest into frame and secure with push buttons.

Folding the Rollator

1. If necessary, remove the accessories and lift the seat.
2. Hold one handle and pull up on the Folding Pull Strap that is attached to the crossbar (locking mechanism) or the pouch underneath the seat.

Using the Hand Brake

Brakes should be used in one of the following methods:
- To Remain Stationary: Hold the hand brake up.
- To Continue Mobility: Release the brake. To unlock, pull up on the top portion of the brake handle.
- To Sit on Seat: Push down on the bottom portion of the brake handle to lock.
- To unlock, pull up on the top portion of the brake handle.
  CAUTION: Always lock hand brakes when user is immobile; never sit on rest seat if hand brakes are not locked.

Adjusting the Height

1. Remove the adjustment knob and bolt by turning counterclockwise.
2. Consult with healthcare professional for proper adjustment.
3. Install the adjustment knob into one of the adjustment holes by turning clockwise.
4. Repeat STEPS 1– 3 for the other side.

Make sure that the height adjustment is at the same height for both handles to ensure safety and proper use.
NOTE: After adjusting the height, the brakes of the rollator must be checked and, if necessary, adjusted.

Attaching Basket, Pouch or Tote Bag

Basket
Install the basket underneath the seat. On some rollator models, the basket is designed to fit underneath the seat or in the front by positioning the hooks on the front seat support frame.

Pouch and Tote Bag
Hang the tote bag between the rollator frame tubes and secure with snap button flaps.

Maintenance

- For dust and dirt, wipe Rollator gently with damp, soft cloth.
- Inspect wheels periodically for wear or damage. Use a mild soap and water solution to clean, rinse with clear water, then dry with a soft cloth.
- Never use organic solvents, abrasive cleaners, or scouring pads on any part of the rollator.
- Before every use, inspect operation of the brakes and have them adjusted if necessary by your GF Health Products, Inc. authorized distributor.
- Inspect rollator weekly or on a set routine preventative maintenance schedule for normal wear and tear. If any component is not in proper working order or appears worn, contact your GF Health Products, Inc. distributor for repair and/or appropriate replacement parts.
- If handgrip or hand brakes are loose, **DO NOT** use. Contact your GF Health Products, Inc. authorized distributor.
- Replace any broken, damaged or worn items immediately.
- For the best service from your rollator, always use replacement parts from GF Health Products, Inc.
- Periodically inspect the casters and caster stems for assembly tightness and verify that the wheels are free of hair or other debris that could interfere with free wheel operation.

Safety Guidelines – Please Read Before Use

1. Intended usage: The rollator has been designed to assist people when walking.
2. Each individual should ALWAYS consult with their physician or other health care professional to determine proper adjustment and usage.
3. The brakes MUST be in the locked position before using the seat.
4. All wheels MUST be in contact with the floor at all times during use. This will ensure the rollator is properly balanced.
5. This product has a maximum user weight limitation of 300 lbs ALWAYS observe the user weight limitation on the hangtag of this product and on the chart referenced at left.
6. Ensure handgrips DO NOT twist on rollator handle - otherwise damage or injury may occur.
7. After installation and before use, ensure that all attaching hardware is securely tightened.
8. ALWAYS use caution when using the rollator on wet and slippery surfaces. DO NOT use the rollator on stairs, steep inclines or uneven terrain.
9. DO NOT hang anything from the frame of the rollator. Items should be placed in the pouch.
10. The rollator basket has a weight limitation of 11 lbs (5 kg). The tote bag has a weight limitation of 10 lbs (5 kg). Items placed in either the basket or the bag should not protrude from the basket or bag.
11. The backrest should ALWAYS be in place and is NOT designed to support the total body weight of the user.
12. Before attempting to reach objects or pick them up from the floor by reaching down between your knees, place feet securely on the floor. Use EXTREME caution when reaching for any object.
13. GF Health Products, Inc. assumes no responsibility for any damage or injury caused by improper installation, assembly or use of this product.
14. Inspect rollator weekly or on a set routine preventative maintenance schedule for normal wear and tear. If any component is not in proper working order or appears worn, contact your Lumex dealer for repair and/or appropriate replacement parts.
15. If the hand brakes fail to stop the Rollator, contact your GF Health Products, Inc. authorized distributor immediately for adjustment or replacement.

WARNING: Indicates a potential hazard situation or unsafe practice that, if not avoided, could result in death or serious personal injury.

CAUTION: Indicates a potential hazard or unsafe practice that, if not avoided, could result in personal injury or product/property damage.

NOTE: Provides application recommendations or other useful information to ensure that you receive the greatest benefit from this product.

WARNING

WARNING: DO NOT use this product without first completely reading and understanding these instructions and any additional instructional material supplied with this product. If you are unable to understand the warnings, cautions or instructions, contact a healthcare professional, dealer or technical personnel before attempting to use this equipment - otherwise, injury or damage may occur.

WARNING: DO NOT use the rollator if there is any damage to the frame.

WARNING: DO NOT use the rollator as a wheelchair or transport device. Rollators are not intended to be propelled while seated.

WARNING: If components are damaged or missing, contact your distributor immediately. DO NOT substitute parts.

WARNING: Before use, ensure that all assembled parts are secure and all hardware is tightened.

WARNING: Care should be taken to ensure that all hand and height adjustments are secure, and that wheels and moving objects are in good working order before using this rollator.

WARNING: When using the rollator in a stationary position, the hand brakes MUST be locked.

WARNING: ALWAYS observe the user weight limitation.

WARNING: Notice for California Customers-
California Proposition 65 WARNING: This product contains a chemical known to the State of California to cause cancer and reproductive or developmental harm.

WARNING: DO NOT attempt to reach objects if you have to move forward in the seat. Reaching for these objects will cause a change to the weight distribution of the rollator and may cause the rollator to tip, resulting in injury or damage.

Inside

Instrucciones para Uso – Lea Todas Las Instrucciones Antes de Uso.

Por favor de consultar con un profesional de cuido de salud para ajustamiento apropiado.

Ensamblaje del Rodador

1. Remueva el Rodador y los otros componentes del cartón.
2. Desdobla el cuadro; extiende las patas delanteras y traseras.
3. Aprieta hacia abajo en el mecanismo de cerrar para asegurar las patas en posición.
   Nota: Asegura que los barras de los lados (mecanismo de cerrar) están posicionados.
4. Inserta mango con manija en el cuadro.
   Nota: Después de instalación, asegure que el cable del freno se asegura al cuadro con el clip.
5. Asegura la posición de los manjas con la perilla de ajustamiento.
6. Instala espaldar en el cuadro y asegura con los botones de apretar.

Doblando el Rodador

1. Si necesario, remueva los accesorios y levanta el asiento.
2. Jale hacia arriba en el Cordón de Doblar que esta fijado en la barra debajo del asiento.

Usando el Freno de Mano

Los frenos deben ser usados en uno de los siguientes métodos:
- Para Mantener estacionario: Aguante el freno hacia arriba.
- Para Continuar Movilidad: Suelta el freno.
- Para Sentarse: Apreté hacia abajo en la porción abajo del freno para cerrar.
- Para Abrir, jale en la porción superior del freno de mano.
  Caución: Siempre cierre los frenos de mano cuando el usuario es inmóvil; nunca se asiente en el asiento si los frenos de mano no están asegurados.

Para Ajustar la Altura

1. Remueve la perilla y tornillo de ajustamiento en sentido contrario a las agujas del reloj.
2. Consulte con un profesional de cuido de salud para ajustamiento apropiado.
3. Instale la perilla de ajustamiento en uno de los huecos en sentido a las agujas del reloj.
4. Repite PASOS 1– 3 para el otro lado.

Asegure que el ajustamiento de altura es lo mismo para los dos mangos para asegurar seguridad y uso apropiado.
Nota: Después de ajustar la altura, los frenos del rodador deben ser examinados y si necesario, ajustados.

Poniendo la Canasta, Bolsa y el Bolsón

Canasta
Instala la canasta debajo del asiento. En algunos de los modelos, la canasta es diseando para poner debajo del asiento o al frente posicionando con los ganchos en la parte del frente del marco.

Bolsa y Bolsón
A gancha el bolsón dentro del marco del rodador y asegura con los solapas.

Mantenimiento

- Para polvo, limpie el rodador con una toalla suave y mojada.
- Examine las ruedas por daño y desgasto periódicamente. Use un jabón suave y agua para limpiar, lave con agua limpia y seca con una toalla suave.
- Nunca use solventes orgánicos, limpiadores abrasivos o estropajos en cual quiere partedel rodador.
- Antes de cada uso, inspección operación de los frenos y permite que su distribuidor de GF Health Products, Inc. se los ajusta si es necesario
- Examine su rodador seminal o en un programa fija de mantenimiento preventiva para función y desgasto. Si algunos componentes so están funcionando correctamente o se ven con desgasto, llame a su distribuidor de GF Health Products, Inc. para reparación y partes de reemplazo.
- Si los mangos o frenos de mano están sueltos, NO LOS use. Llame a su distribuidor autorizado de GF Health Products, Inc.
- Para el mejor servicio de su rodador, siempre use partes de reemplazo de GF Health Products, Inc.
- Periódicamente examine los rueditas y vástagos para su aprieta y verifica que las ruedas están libres de pelo o otros escombros que pueden interferir con el operación libre de la rueda.

Directriz de Seguridad – Por Favor de Leer Antes de Uso

1. Uso Intencionado: Este andador es diseñado para asistir a personas cuando caminando.
2. Cada individual SIEMPRE debe consultar con su doctor o otro profesional de cuido de salud para determinar ajustamiento y uso apropiado.
3. Los frenos TIENEN que estar en la posición asegurados antes de usar el asiento.
4. Todas las ruedas TIENEN que estar en contacto con el piso todo el tiempo cuando en uso. Esto puede asegurar que el rodador esta balanceado apropiadamente.
5. Este producto tiene una limitación de peso máximo de usuario de 300 lb. SIEMPRE tiene que observar la limitación de peso en el "hangtag" de este producto o en la tabla a la izquierda.
6. Asegure que los mangos NO SE dan vuelta en el rodador – o si no, daño o herida puede ocurrir.
7. Después de instalación y antes de uso, asegura que el ferretero esta seguramente apretados.
8. SIEMPRE use caución cuando usando los rodadores en superficies mojados y resbalosos. NO USE el rodador en escaleras, inclinas empinados o terreno disigual.
9. NO fija cual quiere cosa del cuadro del rodador. Cosas deben ser puestos en la bolsa.
10. La canasta del rodador tiene una limitación de peso de 11 lb (5 kg). El bolsón tiene una limitación de 10 lb (5 kg). Cosas puestas en la canasta o bolsón no DEBEN sobresalir.
11. El respaldo debe SIEMPRE estar en posición y no es designado soportar todo el peso del cuerpo del usuario.
12. Antes de atemptar a alcanzar objetos o recojerlos del piso alcanzando dentro sus rodillas, asegura sus pies en el piso. Usa caucion EXTREMO cuando alcanzando por cual quiere objeto.
13. GF Health Products, Inc. no asume ninguin responsabilidad por cual quiere daño o herida causado por instalación, asemblaje o uso incorrecto de este producto.
14. Examine su rodador seminal o en un programa fija de mantenimiento preventiva para función y desgasto. Si algunos componentes no están funcionando y se ve con desgasto, llame a su distribuidor de Lumex para reparación o reemplace de partes.
15. Si los frenos de mano fallan o y parran el rodador, llame a su distribuidor autorizado de GF Health Products, Inc. para ajustamiento o reemplazo.

ADVERTENCIA: Indica un riesgo potencial, situación o practica insegura que si no se evita puede resultar en muerte o herida personal severa.

CAUCIÓN: Indica una amenaza potencial o practica insegura, que si no es evitadapodría resultar en daño personal o a la propiedad.

NOTA: Provee recomendaciones u otra información para asegurar que usted reciba lo mejor de su producto.

ADVERTENCIA

ADVERTENCIA: NO DEBE usar este producto sin leer y entender completamente estas instrucciones y material adicional que viene con este producto. Si no puedes entender los advertencias, cauciones o instrucciones, llame a su profesional, distribuidor o personal técnicos antes de tratar de usar este equipo – o si no, herida o daño puede ocurrir.

ADVERTENCIA: NO DEBE utizar este andador si hay dano al cuadro.

ADVERTENCIA: NO DEBE usar el rodador como una silla de ruedas o silla de transporte. Rodadores son entendidos para propulsado mientras sentado.

ADVERTENCIA: Si los componentes están dañados o si faltan, llame a su distribuidor inmediatamente. NO substituye partes.

ADVERTENCIA: Antes de uso, asegure que las partes instalados están seguros y todo el ferretería esta apretado.

ADVERTENCIA: Cuido debe estar hecho para asegurar que todos los ajustamientos de mano y altura están seguros, y que las ruedas y objetos que se mueven están en buenas condiciones antes de usar el rodador.

ADVERTENCIA: Cuando usando el rodador en la posición estacionado, los frenos de mano TIENEN que estar asegurados.

ADVERTENCIA: SIEMPRE observa la limitación de peso del usuario.

Aviso a Clientes de California
ADVERTENCIA Proposicion 65 de California: Este producto contiene una cemica que el Estado de California de causar cancer y herida reproductiva y de desarrollo.

ADVERTENCIA: NO intente alcanzar para objetos si tiene que moverse hacia adelante en el asiento. Alcanzando para estos objetos puede causar un cambio en la distribución de peso del rodador y puede causar que el rodador se dobla, resultando en herida o daño.

Inside Front Cover | Inside Back Cover

384 mm X 356 mm
Hantag to be printed with 4 Color High Gloss Ink plus UV coating on 80 LB. White Stock.
Red guides are to indicate folds and layout only. They do not print.
Score all folds
Drill holes where indicated, 4mm
Use a CLEAR 25.4 MM round perferated waffer to seal hangtag